



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQ2600.01

TECHNOVIEW® Arixtra CAL 5x1 mL

REF 5090010-RUO

LOT 2611C00.03

🕒 2023-08-31

Inhalt Content	Reagenz Reagent	LOT	🕒
1x1 mL	TECHNOVIEW Arixtra Cal 1	2111C00	2023-08-31
1x1 mL	TECHNOVIEW Arixtra Cal 2	2211C00	2023-08-31
1x1 mL	TECHNOVIEW Arixtra Cal 3	2311C00	2023-08-31
1x1 mL	TECHNOVIEW Arixtra Cal 4	2411C00	2023-08-31
1x1 mL	TECHNOVIEW Arixtra Cal 5	2511C00	2023-08-31

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability Ceveron® alpha	TECHNOVIEW Arixtra Cont low LOT 3U11C00	0.33 – 0.61 µg/mL	0.50 OK
	TECHNOVIEW Arixtra Cont high LOT 3V11C00	0.92 – 1.54 µg/mL	1.26 OK
Etiketten und Barcode Überprüfung Label and bar code Verification		OK	

Durchführung von
Prepared by


Alessandro BENIGNI

05-Jul-2022

Datum / Date:

Freigabe von
Approved by


Lucas HOLBYCZ

05-Jul-2022

Datum / Date: