

CERTIFICATE OF ANALYSIS

CHROMOGENIX werfen

PRODUCT NAME:	Antithrombin 1 x 25 IU	EXPIRATION DATE:	2027-10-31
PRODUCT NUMBER:	00081079639		
LOT NUMBER:	N1057317		

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION
	MIN	MAX			
Antithrombin Activity	25	30	IU/vial	28	PASS
Assay Slope @ 37°C	-0.0024	-0.0034		-0.0028	PASS
Intercept @ 37°C	0.40	0.60		0.49	PASS
r ²	0.990	--		0.999	PASS

We hereby certify that the above product(s):

- have been manufactured and tested according to the requirements of their Manufacturing and Quality Control procedures and have been found to meet all conditions set forth with these documents;

PREPARED BY: Fabiha Bushra *Fabiha Bushra* DATE: 11/4/2025 *11/4/25*
 PRINT SIGNATURE

TITLE: QC Analyst II

APPROVAL SIGNATURE: Nehal Patel *N. M. Patel* DATE: 11/4/2025 *11-4-25*
 PRINT SIGNATURE

TITLE: Senior Group Leader