

CERTIFICATE OF ANALYSIS

CHROMOGENIX **werfen**

PRODUCT NAME: <u>S-2222, 25 mg Substrate</u>	EXPIRATION DATE: <u>2029-06-30</u>
PRODUCT NUMBER: <u>00082031639</u>	
LOT: <u>N0260992</u>	

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION
	MIN	MAX			
SUBSTRATE CONCENTRATION	23.7	26.3	mg/vial	25.0	PASS
FREE p-NITROANILINE	0	0.5	mol %	0.1	PASS

We hereby certify that the above product(s) have been manufactured and tested according to the requirements of their Manufacturing and Quality Control procedures and have been found to meet all conditions set forth with these documents.

PREPARED BY: Brianna Randazzo Brianna Randazzo DATE: 03/23/2026
 PRINT SIGNATURE

TITLE: QC Analyst II

APPROVAL SIGNATURE: Nehal Patel N.M. Patel DATE: 3-23-26
 PRINT SIGNATURE

TITLE: Senior Group Leader