

CERTIFICATE OF ANALYSIS

CHROMOGENIX**werfen**

PRODUCT NAME:	S-2222, 25 mg Substrate	EXPIRATION DATE:	2028-09-30
PRODUCT NUMBER:	00082031639		
LOT:	N0552682		

TEST DESCRIPTION	SPECIFICATIONS			RESULTS	DISPOSITION
	MIN	MAX	UNITS		
SUBSTRATE CONCENTRATION	23.7	26.3	mg/vial	25.4	PASS
FREE p-NITROANILINE	0	0.5	mol %	0	PASS

We hereby certify that the above product(s) have been manufactured and tested according to the requirements of their Manufacturing and Quality Control procedures and have been found to meet all conditions set forth with these documents.

PREPARED BY:

Maria Somoza
PRINTMaria Jomyer
SIGNATURE

DATE: 5/23/2025 05/23/2025

TITLE: QC Analyst II

APPROVAL SIGNATURE:

Nehal Patel
PRINTN. m. Patel
SIGNATURE

DATE: 5/23/2025

5.23.25

TITLE: Senior Group Leader

QCM00820316-2 Rev. 01