

CERTIFICATE OF ANALYSIS

**CHROMOGENIX**

**werfen**

PRODUCT NAME:	S-2238, 25 mg Substrate	EXPIRATION DATE:	2028-09-30
PRODUCT NUMBER:	00082032439		
LOT NUMBER:	N0956138		

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION
	MIN	MAX			PASS/FAIL
SUBSTRATE CONCENTRATION	23.7	26.3	mg/vial	25.7	PASS
FREE p-NITROANILINE	0	0.5	mol %	0	PASS

We hereby certify that the above product(s) have been manufactured and tested according to the requirements of their Manufacturing and Quality Control procedures and have been found to meet all conditions set forth with these documents.

PREPARED BY: Maria Somoza  
PRINT

Maria Somoza  
SIGNATURE

DATE: 9/23/2025 09/23/25

TITLE: QC Analyst II

APPROVAL SIGNATURE: Nehal Patel  
PRINT

N. Patel  
SIGNATURE

DATE: 9/23/2025

9.23.25

TITLE: Senior Group Leader