

CERTIFICATE OF ANALYSIS

CHROMOGENIX**werfen**

PRODUCT NAME:	S-2238, 25 mg Substrate	EXPIRATION DATE:	2028-09-30
PRODUCT NUMBER:	00082032439		
LOT NUMBER:	N0956138		

TEST DESCRIPTION	SPECIFICATIONS			RESULTS	DISPOSITION
	MIN	MAX	UNITS		
SUBSTRATE CONCENTRATION	23.7	26.3	mg/vial	25.7	PASS
FREE p-NITROANILINE	0	0.5	mol %	0	PASS

We hereby certify that the above product(s) have been manufactured and tested according to the requirements of their Manufacturing and Quality Control procedures and have been found to meet all conditions set forth with these documents.

PREPARED BY: Maria Somoza
PRINT*Maria Somoza*
SIGNATUREDATE: 9/23/2025 *09/23/25*

TITLE: QC Analyst II

APPROVAL SIGNATURE: Nehal Patel
PRINT*N. Patel*
SIGNATUREDATE: 9/23/2025 *9.23.25*

TITLE: Senior Group Leader