

CERTIFICATE OF ANALYSIS

CHROMOGENIX

werfen

| | | | |
|-----------------|-------------------------|------------------|------------|
| PRODUCT NAME: | S-2288, 25 mg Substrate | EXPIRATION DATE: | 2028-04-30 |
| PRODUCT NUMBER: | 00082085239 | | |
| LOT NUMBER: | N0451934 | | |

| TEST DESCRIPTION | SPECIFICATIONS | | UNITS | RESULTS | DISPOSITION |
|-------------------------|----------------|------|---------|---------|-------------|
| | MIN | MAX | | | |
| SUBSTRATE CONCENTRATION | 23.7 | 26.3 | mg/vial | 25.3 | PASS |
| FREE p-NITROANILINE | 0 | 0.5 | mol % | 0.5 | PASS |

We hereby certify that the above product(s) have been manufactured and tested according to the requirements of their Manufacturing and Quality Control procedures and have been found to meet all conditions set forth with these documents.

PREPARED BY: Fabiha Bushra
PRINT

Fabiha Bushra
SIGNATURE

DATE: 5/1/2025 5/1/25

TITLE: QC Analyst II

APPROVAL SIGNATURE: Nehal Patel
PRINT

N.M. Patel

SIGNATURE

DATE: 5/1/2025 5-1-25

TITLE: Senior Group Leader