

# CERTIFICATE OF ANALYSIS

**CHROMOGENIX**

**werfen**

PRODUCT NAME:	S-2288, 25 mg Substrate	EXPIRATION DATE: 2028-04-30
PRODUCT NUMBER:	00082085239	
LOT NUMBER:	N0451934	

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION
	MIN	MAX			PASS/FAIL
SUBSTRATE CONCENTRATION	23.7	26.3	mg/vial	25.3	PASS
FREE p-NITROANILINE	0	0.5	mol %	0.5	PASS

We hereby certify that the above product(s) have been manufactured and tested according to the requirements of their Manufacturing and Quality Control procedures and have been found to meet all conditions set forth with these documents.

PREPARED BY: Fabiha Bushra *Fabiha Bushra* DATE: 5/1/2025 5/1/25  
 PRINT SIGNATURE

TITLE: QC Analyst II

APPROVAL SIGNATURE: Nehal Patel *N. M. Patel* DATE: 5/1/2025 5-1-25  
 PRINT SIGNATURE

TITLE: Senior Group Leader