

# CERTIFICATE OF ANALYSIS

**CHROMOGENIX**

**werfen**

|                 |                         |                  |            |
|-----------------|-------------------------|------------------|------------|
| PRODUCT NAME:   | S-2288, 25 mg Substrate | EXPIRATION DATE: | 2028-12-31 |
| PRODUCT NUMBER: | 00082085239             |                  |            |
| LOT NUMBER:     | N1258732                |                  |            |

| TEST DESCRIPTION        | SPECIFICATIONS |      | UNITS   | RESULTS | DISPOSITION |
|-------------------------|----------------|------|---------|---------|-------------|
|                         | MIN            | MAX  |         |         |             |
| SUBSTRATE CONCENTRATION | 23.7           | 26.3 | mg/vial | 24.8    | PASS        |
| FREE p-NITROANILINE     | 0              | 0.5  | mol %   | 0.5     | PASS        |

We hereby certify that the above product(s) have been manufactured and tested according to the requirements of their Manufacturing and Quality Control procedures and have been found to meet all conditions set forth with these documents.

PREPARED BY: Fabiha Bushra  
PRINT

*Fabiha Bushra*  
SIGNATURE

DATE: 1/20/2026 *1/20/26*

TITLE: QC Analyst II

APPROVAL SIGNATURE: Nehal Patel  
PRINT

*N. M. Patel*

SIGNATURE

DATE: 1/20/2026 *1-20-26*

TITLE: Senior Group Leader