

# CERTIFICATE OF ANALYSIS

## CHROMOGENIX werfen

PRODUCT NAME:	FXa, 10 x 71 nkat	EXPIRATION DATE:	2027-09-30
PRODUCT NUMBER:	00082098539		
LOT NUMBER:	N0350651		

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION
	MIN	MAX			
FACTOR Xa, 71 nkat (650515)					PASS/FAIL
ACTIVITY	61	81	nkat/vial	77	PASS
STANDARD CURVE ( $r^2$ )	0.990	1.000	--	1.000	PASS
CONTROL RECOVERY	66.7	73.7	nkat/vial	72.5	PASS

We hereby certify that the above product(s):  
 · have been manufactured and tested according to the requirements of their Manufacturing and Quality Control procedures and have been found to meet all conditions set forth with these documents;

PREPARED BY: Maria Somoza  
 PRINT

Maria Somoza  
 SIGNATURE

DATE: 4/3/2025 04/3/25

TITLE: QC Analyst II

APPROVAL SIGNATURE: Nehal Patel  
 PRINT

N. M. Patel  
 SIGNATURE

DATE: 4-3-25  
4/3/2025

TITLE: Senior Group Leader