



CERTIFICATE OF ANALYSIS

CHROMOGENIX werfen

PRODUCT NAME:	FXa, 10 x 71 nkat	EXPIRATION DATE: 2027-09-30
PRODUCT NUMBER:	00082098539	
LOT NUMBER:	N0350651	

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION
	MIN	MAX			PASS/FAIL
FACTOR Xa, 71 nkat (650515)					
ACTIVITY	61	81	nkat/vial	77	PASS
STANDARD CURVE (r^2)	0.990	1.000	--	1.000	PASS
CONTROL RECOVERY	66.7	73.7	nkat/vial	72.5	PASS

We hereby certify that the above product(s):
 · have been manufactured and tested according to the requirements of their Manufacturing and Quality Control procedures and have been found to meet all conditions set forth with these documents;

PREPARED BY: <u>Maria Somoza</u> PRINT	 SIGNATURE	DATE: <u>4/3/2025</u> <u>04/3/25</u>
TITLE: <u>QC Analyst II</u>		
APPROVAL SIGNATURE: <u>Nehal Patel</u> PRINT	 SIGNATURE	DATE: <u>4-3-25</u> <u>4/3/2025</u>
TITLE: <u>Senior Group Leader</u>		