

CERTIFICATE OF ANALYSIS

CHROMOGENIX **werfen**

PRODUCT NAME:	FXa, 10 x 71 nkat	EXPIRATION DATE:	2028-09-30
PRODUCT NUMBER:	00082098539		
LOT NUMBER:	N0361767		

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION
	MIN	MAX			
FACTOR Xa, 71 nkat (650515)					PASS/FAIL
ACTIVITY	61	81	nkcat/vial	76	PASS
STANDARD CURVE (r^2)	0.990	1.000	--	1.000	PASS
CONTROL RECOVERY	66.7	73.7	nkcat/vial	72.2	PASS

We hereby certify that the above product(s):

· have been manufactured and tested according to the requirements of their Manufacturing and Quality Control procedures and have been found to meet all conditions set forth with these documents;

PREPARED BY: Monika Skowronska
PRINT

Monika Skowronska
SIGNATURE

DATE: 4-10-26
4/10/2026

TITLE: QC Analyst III

APPROVAL SIGNATURE: Nehal Patel
PRINT

N.M. Patel

SIGNATURE

DATE: 4-10-26
4/10/2026

TITLE: Senior Group Leader