

CERTIFICATE OF ANALYSIS

# CHROMOGENIX werfen

PRODUCT NAME:	<u>FXa, 10 x 71 nkat</u>	EXPIRATION DATE:	<u>2028-05-31</u>
PRODUCT NUMBER:	<u>00082098539</u>		
LOT NUMBER:	<u>N1158150</u>		

TEST DESCRIPTION	SPECIFICATIONS			RESULTS	DISPOSITION
	MIN	MAX	UNITS		
FACTOR Xa, 71 nkat (650515)					PASS/FAIL
ACTIVITY	61	81	nkat/vial	78	PASS
STANDARD CURVE (r <sup>2</sup> )	0.990	1.000	--	1.000	PASS
CONTROL RECOVERY	66.7	73.7	nkat/vial	72.7	PASS

We hereby certify that the above product(s):  
 · have been manufactured and tested according to the requirements of their Manufacturing and Quality Control procedures and have been found to meet all conditions set forth with these documents;

PREPARED BY: Maria Somoza Maria Somoza  
 PRINT SIGNATURE

DATE: 12/2/2025 12/2/25

TITLE: QC Analyst II

APPROVAL SIGNATURE: Nehal Patel N.M. Patel  
 PRINT SIGNATURE

DATE: 12/2/2025 12.2.25

TITLE: Senior Group Leader