CHROMOGENIX

werfen

EXPIRATION DATE:	2028-03-31

TEST DESCRIPTION	SPECIFICATIONS			RESULTS	DISPOSITION
	MIN	MAX	UNITS		PASS/FAIL
SUBSTRATE CONCENTRATION	23.7	26.3	mg/vial	25.0	PASS
FREE p-NITROANILINE	0	0.5	mol %	0.2	PASS

We hereby certify that the above product(s) have been manufactured and tested according to the requirements of their Manufacturing and Quality Control procedures and have been found to meet all conditions set forth with these documents.

PREPARED BY:	Fabiha Bushra PRINT	Fabi ha Bush pa	DATE:	4/7/2025 4/7/25
TIT	LE: 20 Analyst 99	SIGNATURE		4-7:25
APPROVAL SIGNATU	RE: Nehal Patel PRINT	12 m. Putel SIGNATURE	DATE:	4/7/2025
TIT	l.E: Senior Group Leader			