## **CERTIFICATE OF ANALYSIS**

## CHROMOGENIX werfen

PRODUCT NAME:	COATEST APC Resistance Control Plasma Level 1	EXPIRATION DATE: 2027-04-30		
PRODUCT NUMBER:	00082265063			
LOT	N0351418	_		

Analysis with COATEST APC R V:						
APT Time	<u>CaCl₂</u>		39	seconds	34	PASS
	APC/CaCl <sub>2</sub>	75		seconds	101	PASS
APC Ratio		2.7	3.2		2.9	PASS

We hereby certify that the above product(s):

 have been manufactured and tested according to the requirements of their Manufacturing and Quality Control procedures and have been found to meet all conditions set forth with these documents.

PREPARED BY:	PRINT	Monika Skowronska	Monika SIGNATI	Skomronska URE	DATE:	4/29/2025	4-29-25
TITLE:		QC Analyst II					
APPROVAL SIGNATURE:	PRINT	Nehal Patel	p.m. Pa	-	DATE:	4/29/2025	H-29.25
TITLE:	FINIT	Senior Group Leader	SIGNATI	UKE			