

CERTIFICATE OF ANALYSIS

CHROMOGENIX *werfen*

PRODUCT NAME:	COATEST APC Resistance Control Plasma Level 1	EXPIRATION DATE:	2027-04-30
PRODUCT NUMBER:	00082265063		
LOT	N0351418		

Analysis with COATEST APC R V:						
APT Time	<u>CaCl₂</u>	--	39	seconds	34	PASS
	<u>APC/CaCl₂</u>	75	--	seconds	101	PASS
APC Ratio		2.7	3.2	--	2.9	PASS

We hereby certify that the above product(s):

- have been manufactured and tested according to the requirements of their Manufacturing and Quality Control procedures and have been found to meet all conditions set forth with these documents.

PREPARED BY: Monika Skowronska *Monika Skowronska* DATE: 4/29/2025 *4-29-25*
 PRINT SIGNATURE

TITLE: QC Analyst II

APPROVAL SIGNATURE: Nehal Patel *N.M. Patel* DATE: 4/29/2025 *4-29-25*
 PRINT SIGNATURE

TITLE: Senior Group Leader