CERTIFICATE OF ANALYSIS

CHROMOGENIX werfen

 PRODUCT NAME:
 COATEST APC Resistance Control Plasma Level 1
 EXPIRATION DATE: 2027-04-30

 PRODUCT NUMBER:
 00082265063
 N0351418

Analysis with COATEST APC R V:						
APT Time	CaCl ₂		39	seconds	34	PASS
	APC/CaCl ₂	75		seconds	101	PASS
APC Ratio		2.7	3.2		2.9	PASS

We hereby certify that the above product(s):

- have been manufactured and tested according to the requirements of their Manufacturing and Quality Control procedures and have been found to meet all conditions set forth with these documents.

PREPARED BY:	Monika Skowronska	Monika SIGNAT	Skowronska	DATE:	4/29/2025	4-29-25
TITLE:	QC Analyst II					
APPROVAL SIGNATURE: PF	Nehal Patel	p.m. Pa		DATE:	4/29/2025	4.29.25
TITLE:	Senior Group Leader					