

CERTIFICATE OF ANALYSIS

CHROMOGENIX *werfen*

PRODUCT NAME:	COATEST APC Resistance Control Plasma Level 2	EXPIRATION DATE:	2027-02-28
PRODUCT NUMBER:	00082266863		
LOT NUMBER:	N0250218		

TEST DESCRIPTION	SPECIFICATIONS			RESULTS	DISPOSITION
	MIN	MAX	UNITS		
Analysis with COATEST APC R V:					PASS/FAIL
APT Time <u>CaCl₂</u>	28	41	seconds	37	PASS
APC Ratio	1.4	1.9	--	1.9	PASS

We hereby certify that the above product(s):
 · have been manufactured and tested according to the requirements of their Manufacturing and Quality Control procedures and have been found to meet all conditions set forth with these documents;

PREPARED BY: Monika Skowronska *Monika Skowronska* DATE: 3/25/2025 *3-25-25*
 PRINT SIGNATURE

TITLE: QC Analyst II

APPROVAL SIGNATURE: Nehal Patel *N.m. Patel* DATE: 3/25/2025 *3-25-25*
 PRINT SIGNATURE

TITLE: Senior Group Leader