## **CERTIFICATE OF ANALYSIS**

## CHROMOGENIX werfen

PRODUCT NAME:	COATEST APC Resistance Control Plasma Level 2	EXPIRATION DATE: 2027-02-28	
PRODUCT NUMBER:	00082266863		
LOT NUMBER:	N0250218		

TEST DESCRIPTION	SPECIFICATIONS			RESULTS	DISPOSITION
Analysis with COATEST APC R V:	MIN	MAX	UNITS		PASS/FAIL
APT Time <u>CaCl</u> <sub>2</sub>	28	41	seconds	37	PASS
APC Ratio 2255 (ASSAULT) A STATE OF THE STAT	1.4	1.9		1.9	PASS

PREPARED BY:	Monika Skowronska PRINT	Honika Skopromka . SIGNATURE	DATE: _	3/25/2025 3-25-25
TITLE:	QC Analyst II			
APPROVAL SIGNATURE:	Nehal Patel	N.m. Pater	DATE:	3-25-25-
	PRINT	SIGNATURE		n realization of the second of
TITLE:	Senior Group Lea	nder		

We hereby certify that the above product(s):

have been manufactured and tested according to the requirements of their Manufacturing and Quality Control procedures and have been found to meet all conditions set forth with these documents;