

CERTIFICATE OF ANALYSIS

CHROMOGENIX **werfen**

PRODUCT NAME:	COATEST APC Resistance Control Plasma Level 2	EXPIRATION DATE: 2027-02-28
PRODUCT NUMBER:	00082266863	
LOT NUMBER:	N0250218	

TEST DESCRIPTION	SPECIFICATIONS	RESULTS	DISPOSITION
	MIN	MAX	PASS/FAIL
Analysis with COATEST APC R V:		UNITS	
APT Time <u>CaCl₂</u>	28	41 seconds	PASS
APC Ratio	1.4	1.9 --	PASS

We hereby certify that the above product(s):
 · have been manufactured and tested according to the requirements of their Manufacturing and Quality Control procedures and have been found to meet all conditions set forth with these documents;

PREPARED BY: <u>Monika Skowronska</u> <i>Monika Skowronska</i> PRINT SIGNATURE	DATE: <u>3/25/2025</u> <i>3-25-25</i>
TITLE: <u>QC Analyst II</u>	
APPROVAL SIGNATURE: <u>Nehal Patel</u> <i>N. M. Patel</i> PRINT SIGNATURE	DATE: <u>3/25/2025</u> <i>3-25-25</i>
TITLE: <u>Senior Group Leader</u>	