CERTIFICATE OF ANALYSIS

CHROMOGENIX werfen

PRODUCT NAME:	COATEST APC Resistance Control Plasma Level 2	EXPIRATION DATE: 2027-02-28
PRODUCT NUMBER:	00082266863	
LOT NUMBER:	N0250218	

TEST DESCRIPTION		SPECIFICATIONS			RESULTS	DISPOSITION
Analysis with COATEST APC R V:		MIN	MAX	UNITS		PASS/FAIL
APT Time	$CaCl_2$	28	41	seconds	37	PASS
APC Ratio		1.4	1.9	--	1.9	PASS

We hereby certify that the above product(s):
· have been manufactured and tested according to the requirements of their Manufacturing and Quality Control procedures and have been found to meet all conditions set forth with these documents;

PREPARED BY: Monika Skowronska (PRINT) Monika Skowronska (SIGNATURE) DATE: 3/25/2025 3-25-25

TITLE: QC Analyst II

APPROVAL SIGNATURE: Nehal Patel (PRINT) N.M. Patel (SIGNATURE) DATE: 3/25/2025 3.25.25

TITLE: Senior Group Leader

526 Route 303, Orangeburg, New York 10962, Telephone 845-365-8000