

CERTIFICATE OF ANALYSIS

CHROMOGENIX **werfen**

PRODUCT NAME:	COATEST APC Resistance Control Plasma Level 2	EXPIRATION DATE: 2028-06-30
PRODUCT NUMBER:	00082266863	
LOT NUMBER:	N0664644	

TEST DESCRIPTION	SPECIFICATIONS			RESULTS	DISPOSITION
Analysis with COATEST APC R V:	MIN	MAX	UNITS		PASS/FAIL
APT Time <u>CaCl₂</u>	28	41	seconds	35	PASS
APC Ratio	1.4	1.9	--	1.8	PASS

We hereby certify that the above product(s):
 · have been manufactured and tested according to the requirements of their Manufacturing and Quality Control procedures and have been found to meet all conditions set forth with these documents;

PREPARED BY: Monika Skowronska Monika Skowronska DATE: 7/9/2026 7-9-26
 PRINT SIGNATURE

TITLE: QC Analyst III

APPROVAL SIGNATURE: Nehal Patel N.M. Patel DATE: 7/9/2026 7-9-26
 PRINT SIGNATURE

TITLE: Senior Group Leader