

CERTIFICATE OF ANALYSIS



PRODUCT NAME:	COATEST APC Resistance V	EXPIRATION DATE:	2026-09-30
PRODUCT NUMBER:	00082312063		
LOT NUMBER:	N0351161		

TEST DESCRIPTION	SPECIFICATIONS			RESULTS	Pass/Fail
	MIN	MAX	UNITS		
Diluted Control Plasma Level 1	28	44	seconds	32	PASS
Diluted Control Plasma Level 2	30	50	seconds	34	PASS

APC-V ratio on ACL				--	--
	Control Plasma Level 1	2.8	3.2		
Control Plasma Level 2		1.4	1.9	--	1.9
					PASS

KIT COMPONENTS:

NAME	LOT NUMBER	VOLUME
COATEST APC Resistance CaCl ₂	N0350849	8 mL
COATEST APC Resistance APTT reagent	N0350844	16 mL
COATEST APC Resistance APC/CaCl ₂	N0351157	2 mL
COATEST APC Resistance Control Plasma Level 1	N0351160	1 mL
COATEST APC Resistance Control Plasma Level 2	N0250218	1 mL
COATEST APC Resistance V-DEF Plasma	N0350657	4 mL

We hereby certify that the above product(s):

- have been manufactured and tested according to the requirements of their Manufacturing and Quality Control procedures and have been found to meet all conditions set forth with these documents.

PREPARED BY: Monika Skowronska DATE: 4/17/2025
 PRINT Monika Skowronska SIGNATURE Monika Skowronska

4-17-25

APPROVAL SIGNATURE: Nehal Patel DATE: 4/17/2025
 PRINT Nehal Patel SIGNATURE Nehal Patel

4-17-25

TITLE: Senior Group Leader