

CERTIFICATE OF ANALYSIS



PRODUCT NAME:	COATEST APC Resistance V-S	EXPIRATION DATE:	2027-06-30
PRODUCT NUMBER:	00082313863		
LOT:	N0160244		

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	Pass/Fail
	MIN	MAX			
Diluted Control Plasma Level 1	27	43	seconds	35	PASS
Diluted Control Plasma Level 2	28	48	seconds	36	PASS

APC-V ratio on ACL	MIN	MAX	UNITS	RESULTS	Pass/Fail
Control Plasma Level 1	2.8	3.2	--	3.0	PASS
Control Plasma Level 2	1.4	1.9	--	1.8	PASS

KIT COMPONENTS:

NAME	LOT NUMBER	VOLUME
COATEST APC Resistance CaCl2	N1258726	2 mL
COATEST APC Resistance APTT reagent	N1258728	4 mL
COATEST APC Resistance APC/CaCl2	N0169890	2 mL
COATEST APC Resistance Control Plasma Level 1	N0169897	1 mL
COATEST APC Resistance Control Plasma Level 2	N0169611	1 mL
COATEST APC Resistance V-DEF Plasma	N1259082	4 mL

We hereby certify that the above product(s):

- have been manufactured and tested according to the requirements of their Manufacturing and Quality Control procedures and have been found to meet all conditions set forth with these documents.

PREPARED BY: Monika Skowronska *Monika Skowronska* DATE: 2-9-26
PRINT SIGNATURE 2/9/2026

TITLE: QC Analyst II

APPROVAL SIGNATURE: Nehal Patel *N.M. Patel* DATE: 2/9/2026
PRINT SIGNATURE 2/9/2026

TITLE: Senior Group Leader