

CERTIFICATE OF ANALYSIS

**CHROMOGENIX** werfen

PRODUCT NAME:	COAMATIC HEPARIN	EXP DATE:	2027-03-31
PRODUCT NUMBER:	00082339363		
LOT NUMBER:	N0351183		

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION
	MIN	MAX			
<b>A. FACTOR Xa</b>					
ACTIVITY	31.5	38.5	nkat/vial	37.6	PASS
STANDARD CURVE (r <sup>2</sup> )	0.990	1.000	--	1.000	PASS
CONTROL RECOVERY	66.7	73.7	nkat/vial	70.3	PASS

<b>B. S-2732, FOR COAMATIC HEPARIN</b>					
SUBSTRATE CONCENTRATION	14.3	15.7	mg/vial	14.4	PASS
FREE p-NITROANILINE	0	0.5	mol %	0.1	PASS

<b>C. COAMATIC HEPARIN KIT - LMW HEPARIN ON TOP</b>					
SLOPE	-0.351	-0.103	--	-0.191	PASS
r <sup>2</sup>	0.985	--	--	0.997	PASS

<b>D. COAMATIC HEPARIN KIT - UNFRACTIONATED HEPARIN ON TOP</b>					
SLOPE	-0.336	-0.114	--	-0.193	PASS
r <sup>2</sup>	0.985	--	--	0.998	PASS

**KIT COMPONENTS:**

NAME	LOT NUMBER	VOLUME
COAMATIC HEPARIN FACTOR Xa	N0351180	1 mL
COAMATIC HEPARIN S-2732	N0351181	2 mL

We hereby certify that the above product(s):  
 · have been manufactured and tested according to the requirements of their Manufacturing and Quality Control procedures and have been found to meet all conditions set forth with these documents;

PREPARED BY: Brianna Randazzo *Brianna Randazzo* DATE: 5/7/2025  
 PRINT SIGNATURE

TITLE: QC Analyst II

APPROVAL SIGNATURE: Nehal Patel *N.M. Patel* DATE: 5-7-25  
 PRINT SIGNATURE

TITLE: Senior Group Leader