

CERTIFICATE OF ANALYSIS

CHROMOGENIX

werfen

PRODUCT NAME: <u>COATEST SP FVIII</u>	EXPIRATION DATE: <u>2027-08-31</u>
PRODUCT NUMBER: <u>00082408663</u>	
LOT NUMBER: <u>N0160090</u>	

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION
	MIN	MAX			
High Range					
Slope	0.0077	0.0112	--	0.0088	PASS
Intercept	--	0.250	--	0.074	PASS
Standard Curve (r)	0.99	--	--	1.000	PASS
Low Range					
Slope	0.020	0.060	--	0.042	PASS
Intercept	--	0.406	--	0.082	PASS
Standard Curve (r)	0.99	--	--	0.996	PASS

KIT COMPONENTS:

NAME	LOT NUMBER	VOLUME
COATEST SP FVIII CaCl2	N0855805	6 mL
COATEST SP FVIII Factor IXa + Factor X	N1057580	10 mL
COATEST SP FVIII S-2765 + I-2581	N0956137	12 mL
COATEST SP FVIII Buffer, stock solution	N0855546	20 mL
COATEST SP FVIII Phospholipid	N0160089	2 mL

We hereby certify that the above product(s):

· have been manufactured and tested according to the requirements of their Manufacturing and Quality Control procedures and have been found to meet all conditions set forth with these documents.

PREPARED BY: Monika Skowronska *Monika Skowronska* DATE: 2-24-26
 PRINT SIGNATURE 02/24/26

TITLE QC Analyst II

APPROVAL SIGNATURE: Nehal Patel *N.M. Patel* DATE: 2-24-26
 PRINT SIGNATURE 02/24/26

TITLE Senior Group Leader