CERTIFICATE OF ANALYSIS

CHROMOGENIX werfen

PRODUCT NAME:	COATEST APC Resistance Control Plasma Level 1	EXPIRATION DATE: 2027-04-30		
PRODUCT NUMBER:	00082265063			
LOT	N0351418			

Analysis with COAT	EST APC R V:					
APT Time	CaCl ₂		39	seconds	34	PASS
* * :	APC/CaCl ₂	75		seconds	101	PASS
APC Ratio	D 3	2.7	3.2		2.9	PASS

We hereby certify that the above product(s):

- have been manufactured and tested according to the requirements of their Manufacturing and Quality Control procedures and have been found to meet all conditions set forth with these documents.

PREPARED BY:PRINT	Monika Skowronska	Monika Sko	uronska	DATE:	4/29/2025	4-29-25	5
TITLE:	QC Analyst II	SIGNATURE					
APPROVAL SIGNATURE:	Nehal Patel	p.m. Padel		DATE:	4/29/2025	H : 29.2	
PRINT	7.4% 2.6%	SIGNATURE	(11)	Kt) to your min	1 MANAGAMENT TO	23,97,057,7636 : -40-117,506	
TITLE:	Senior Group Leader	-	and a section of the			27.07.01.0.207C	