

CERTIFICATE OF ANALYSIS

CHROMOGENIX



PRODUCT NAME:	COATEST APC Resistance Control Plasma Level 1	EXPIRATION DATE: 2026-03
PRODUCT NUMBER:	00082265063	
LOT	N0341018	

Analysis with COATEST APC R V:						
APT Time	CaCl ₂	--	39	seconds	33	PASS
	APC/CaCl ₂	75	--	seconds	90	PASS
APC Ratio		2.7	3.2	--	2.7	PASS

We hereby certify that the above product(s):

- have been manufactured and tested according to the requirements of their Manufacturing and Quality Control procedures and have been found to meet all conditions set forth with these documents;
- meet all applicable requirements of the 98/79/EC Directive according to which they are CE marked.

PREPARED BY: Monika Skowronska Monika Skowronska DATE: 4/23/2024 4-23-24
 PRINT SIGNATURE

TITLE: QC Analyst II

APPROVAL SIGNATURE: Nehal Patel N.m. Patel DATE: 4/23/2024 4-23-24
 PRINT SIGNATURE

TITLE: Senior QC Analyst