

CERTIFICATE OF ANALYSIS

CHROMOGENIX werfen

PRODUCT NAME:	COATEST APC Resistance Control Plasma Level 2	EXPIRATION DATE: 2027-02-28
PRODUCT NUMBER:	00082266863	
LOT NUMBER:	N0250218	

TEST DESCRIPTION	SPECIFICATIONS		RESULTS	DISPOSITION
Analysis with COATEST APC R V:	MIN	MAX	UNITS	PASS/FAIL
APT Time <u>CaCl₂</u>	28	41	seconds	PASS
APC Ratio	1.4	1.9	1.9	PASS

We hereby certify that the above product(s):
 · have been manufactured and tested according to the requirements of their Manufacturing and Quality Control procedures and have been found to meet all conditions set forth with these documents;

PREPARED BY: Monika Skowronska Monika Skowronska DATE: 3/25/2025 3-25-25
 PRINT SIGNATURE

TITLE: QC Analyst II

APPROVAL SIGNATURE: Nehal Patel N.m. Patel DATE: 3/25/2025 3-25-25
 PRINT SIGNATURE

TITLE: Senior Group Leader