

M30 Apoptosense® ELISA

REF 10011

Kit Lot No. LOT 10011-024	Expiry  2026-06
--	--

Component		A _{450 nm}	LOT	
M30 Standard A	0 U/L	0,105	10011SA-024	2026-07
M30 Standard B	75 U/L	0,241	10011SB-024	2026-07
M30 Standard C	150 U/L	0,415	10011SC-024	2026-07
M30 Standard D	250 U/L	0,689	10011SD-024	2026-07
M30 Standard E	500 U/L	1,480	10011SE-024	2026-07
M30 Standard F	750 U/L	2,242	10011SF-024	2026-07
M30 Standard G	1000 U/L	2,804	10011SG-024	2026-07
M5 Coated Microstrips			M5-029	2028-06
M30 HRP Conjugate			10011C-024	2026-06
M30 Conjugate Dilution Buffer			10011CD-024	2026-06
TMB Substrate			VLVT-012	2029-03
Stop Solution			VLVSS-021	2028-03
Wash Tablet			VLVWT-012	2027-08
Controls	Expected value (U/L)	Obtained value (U/L)	LOT	
M30 Control Low	125 ± 25 U/L	119,82U/L	10011CL-024	2026-07
M30 Control High	650 ± 100 U/L	646,30U/L	10011CH-024	2026-07

Test date: 2025-08-26	Sign: <i>Hoszaip</i>
-----------------------	----------------------