

# M30 Apoptosense® ELISA

REF 10011

|                                        |                                                                                                        |
|----------------------------------------|--------------------------------------------------------------------------------------------------------|
| Kit Lot No.<br><b>LOT</b><br>10011-026 | Expiry<br><br>2027-04 |
|----------------------------------------|--------------------------------------------------------------------------------------------------------|

| Component                     |                      | A <sub>450</sub> nm  | <b>LOT</b> |     |    |
|-------------------------------|----------------------|----------------------|------------|-----|---------------------------------------------------------------------------------------|
| M30 Standard A                | 0 U/L                | 0.121                | 10011SA-   | 026 | 2027-06                                                                               |
| M30 Standard B                | 75 U/L               | 0.215                | 10011SB-   | 026 | 2027-06                                                                               |
| M30 Standard C                | 150 U/L              | 0.323                | 10011SC-   | 026 | 2027-06                                                                               |
| M30 Standard D                | 250 U/L              | 0.578                | 10011SD-   | 026 | 2027-06                                                                               |
| M30 Standard E                | 500 U/L              | 1.250                | 10011SE-   | 026 | 2027-06                                                                               |
| M30 Standard F                | 750 U/L              | 1.994                | 10011SF-   | 026 | 2027-06                                                                               |
| M30 Standard G                | 1000 U/L             | 2.684                | 10011SG-   | 026 | 2027-06                                                                               |
| M5 Coated Microstrips         |                      |                      | M5-        | 031 | 2029-05                                                                               |
| M30 HRP Conjugate             |                      |                      | 10011C-    | 026 | 2027-04                                                                               |
| M30 Conjugate Dilution Buffer |                      |                      | 10011CD-   | 026 | 2027-04                                                                               |
| TMB Substrate                 |                      |                      | VLVT-      | 012 | 2029-03                                                                               |
| Stop Solution                 |                      |                      | VLVSS-     | 022 | 2029-01                                                                               |
| Wash Tablet                   |                      |                      | VLVWT-     | 013 | 2028-09                                                                               |
| Controls                      | Expected value (U/L) | Obtained value (U/L) | <b>LOT</b> |     |  |
| M30 Control Low               | 125 ± 25 U/L         | 120.32U/L            | 10011CL-   | 026 | 2027-06                                                                               |
| M30 Control High              | 850 ± 100 U/L        | 624.03U/L            | 10011CH-   | 026 | 2027-06                                                                               |

|                       |                                                                                           |
|-----------------------|-------------------------------------------------------------------------------------------|
| Test date: 2026-07-13 | Sign:  |
|-----------------------|-------------------------------------------------------------------------------------------|