



# Certificate of Analysis

## Factor Xla Control

<b>REF</b>	Catalog number	1188
<b>LOT</b>	Batch code	HHW1934
	Use by	2029 08
	Upper limit of temperature	-20°C

Activity after reconstitution with 4.0 mL water:

**Acceptance range**  
**8.1 – 12.1 mIU/mL**

The above Factor Xla activity apply only to Lot#HHW1934

The calibration was performed using the Rox Factor Xla kit (Ref#110050). Factor Xla Calibrator (Ref#1199), potency assigned vs the 1<sup>st</sup> International standard for Activated Blood Coagulation Factor XI (FXIa), Human, NIBSC code 13/100, was used as calibrator.

This Lot# has been approved for release after analyses according to established protocols with results demonstrating fulfillment of product specifications. All data concerning materials used, production data and release testing are documented and traceable under the individual Lot# which is stated on each package.



Pia Bryngelhed  
Quality Manager, Rossix AB

September 5, 2024

Date



Rossix AB  
Taljegardsgatan 3B  
SE-431 53 Molndal  
Sweden

Phone: +46 31 706 8965  
[info@rossix.com](mailto:info@rossix.com)  
[www.rossix.com](http://www.rossix.com)