

# CERTIFICATE OF ANALYSIS TECHNOPLASTIN HIS Ceveron



Attachment 10 DQ6T00.05

6T42C.02

Temp.: TF0256.07

REF 5003010-RUO  
LOT 6T42C.02  
2026-08-31

Analytical Procedures		Requirement		Result	
visual inspection	Cap color	blue		confirmed	OK
	Vial type	clear glass vial		confirmed	OK
	Vial label	according to packaging SOP		confirmed	OK
	Vial barcode	according to packaging SOP		confirmed	OK
	Box label	according to packaging SOP		confirmed	OK
	Box barcode	according to packaging SOP		confirmed	OK
	Package insert	present		confirmed	OK
	Batch table	present		confirmed	OK
	Appearance before reconstitution	white colored plug of lyophilized material		confirmed	OK
	Appearance after reconstitution	turbid liquid		confirmed	OK
Recovery of controls	<b>Control N</b> LOT 1P31C01	0.83	- 1.11 INR	0.91 INR	OK
	<b>Control A</b> LOT 3P42C00	2.02	- 2.72 INR	2.48 INR	OK

Prepared by

Michael KAFKA

25. OKT. 2024

Date

Approved by

Lucas HOŁOWICZ

25. OKT. 2024

Date