

**CERTIFICATE OF ANALYSIS
TECHNOPLASTIN HIS Ceveron**



Attachment 10 DQ6T00.05

Temp.: TP624.07

6T42C.02

REF	5003010-RUO
LOT	6T42C.02
	2026-08-31

Analytical Procedures	Requirement	Result
visual inspection	Cap color	blue confirmed OK
	Vial type	clear glass vial confirmed OK
	Vial label	according to packaging SOP confirmed OK
	Vial barcode	according to packaging SOP confirmed OK
	Box label	according to packaging SOP confirmed OK
	Box barcode	according to packaging SOP confirmed OK
	Package insert	present confirmed OK
	Batch table	present confirmed OK
	Appearance before reconstitution	white colored plug of lyophilized material confirmed OK
	Appearance after reconstitution	turbid liquid confirmed OK
Recovery of controls	Control N LOT 1P31C01	0.83 - 1.11 INR 0.91 INR OK
	Control A LOT 3P42C00	2.02 - 2.72 INR 2.48 INR OK

Prepared by

Michael KAFKA

25. OKT. 2024

Date

Approved by

Lucas HODOVICZ

25. OKT. 2024

Date