

**CERTIFICATE OF ANALYSIS  
TECHNOPLASTIN HIS Ceveron**



Attachment 10      DQ6T00.05

6T52C.02

Temp.: TFC326.07

REF      5003010-RUO  
 LOT      6T52C.02  
      2027-06-30

Analytical Procedures		Requirement	Result	
visual inspection	Cap color	blue	confirmed	OK
	Vial type	clear glass vial	confirmed	OK
	Vial label	according to packaging SOP	confirmed	OK
	Vial barcode	according to packaging SOP	confirmed	OK
	Box label	according to packaging SOP	confirmed	OK
	Box barcode	according to packaging SOP	confirmed	OK
	Package insert	present	confirmed	OK
	Batch table	present	confirmed	OK
	Appearance before reconstitution	white colored plug of lyophilized material	confirmed	OK
	Appearance after reconstitution	turbid liquid	confirmed	OK
Recovery of controls	<b>Control N</b> LOT 1P42C01	0.79 - 1.07 INR	0.99 INR	OK
	<b>Control A</b> LOT 3P42C01	2.02 - 2.72 INR	2.35 INR	OK

Prepared by

Lavdrim NASUFI

24. FEB. 2026

Date

Approved by

Lukas GRUBELNIG

24. FEB. 2026

Date