

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ0T00AZ
--	---	----------

TECHNOTHROMBIN TGA Kit 3x16 T.

REF 5006010
LOT 0T51C00.01
 2026-04-30

Inhalt Content	Reagenz Reagent	LOT	
1x3mL	TGA CAL BUF	9P51C00	2027-03-31
3x1.5mL	TGA SUB	8F51C00	2027-02-28
1x0.5mL	TGA RC Low	7D51C00	2027-03-31
1x0.5mL	TGA RC High	7F53C00	2027-02-28
1x1.5mL	TGA RD	7G51C00	2027-03-31
1x0.5mL	TGA CAL	5T41C	2026-06-30
1x1mL	TGA Control High	1Z53C00	2026-05-31
1x1mL	TGA Control Low	2852C00	2026-04-30

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Kontrollwiederfindung RC Low	TGA Control High (nM Peak Thrombin) LOT 1Z53C00	312 - 520	470	OK
Recovery of controls RC Low	TGA Control Low (nM Peak Thrombin) LOT 2852C00	121 - 299	275	OK
Kontrollwiederfindung RC High	TGA Control High (nM Peak Thrombin) LOT 1Z53C00	422 - 704	644	OK
Recovery of controls RC High	TGA Control Low (nM Peak Thrombin) LOT 2852C00	240 - 416	403	OK
Kontrollwiederfindung RD	TGA Control High (nM Peak Thrombin) LOT 1Z53C00	410 - 683	592	OK
Recovery of controls RD	TGA Control Low (nM Peak Thrombin) LOT 2852C00	404 - 673	547	OK

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ0T00AZ
---	---	-----------------

Durchführung von
Prepared by

Fikret NASUFI *Fikret Nasufi*

08. JULI 2025

Datum / Date:

Freigabe von
Approved by

Lucas HOŁOVICZ

08. JULI 2025

Datum / Date: