

	<b>ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS</b>	DQB300AZ
---	---	----------

## Ceveron® TGA RB Kit 180T

**REF**      5006011-RUO  
**LOT**      B342C00.02  
      2025-05-31

Inhalt Content	Reagenz Reagent	<b>LOT</b>	
3x1 mL	Ceveron® TGA RB	7T31C	2025-05-31
3x1.5 mL	Ceveron® TGA BUF	9Q41C	2026-09-30
3x3 mL	Ceveron® TGA SUB	8E41C00	2026-03-31
3x1 mL	Ceveron® TGA Control High	1Y42C00	2025-08-31
3x1 mL	Ceveron® TGA Control Low	2742C00	2025-08-31
1x25 mL	Ca-Chloridlösung 25mmol/L	9C41CB0	2029-03-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Kontrollwiederfindung Recovery of controls	Ceveron® TGA Control-High (nM Peak Thrombin) <b>LOT</b> 1Y42C00	107 - 178	157	OK
	Ceveron® TGA Control-Low (nM Peak Thrombin) <b>LOT</b> 2742C00	61 - 101	88	OK
Etiketten und Barcode Überprüfung Label and bar code verification			OK	

**Durchführung von  
Prepared by**

Fikret NASUFI *Fikret Nasufi*

21. OKT. 2024

Datum / Date:

**Freigabe von  
Approved by**

Lucas H. *Lucas H. Wicz*

21. OKT. 2024

Datum / Date: