



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQB300AZ

Ceveron TGA RB Kit 180T

REF 5006011-RUO

LOT B351C00.02

2026-01-31

Inhalt Content	Reagenz Reagent	LOT	
3x1 mL	Ceveron® TGA RB	7T41C	2026-09-30
3x1.5 mL	Ceveron® TGA BUF	9Q41C	2026-09-30
3x3 mL	Ceveron® TGA SUB	8E41C00	2026-03-31
3x1 mL	Ceveron® TGA Control High	1Y51C00	2026-01-31
3x1 mL	Ceveron® TGA Control Low	2751C00	2026-01-31
1x25 mL	Ca-Chloridlösung 25mmol/L	9C41CB0	2029-03-31

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result	
Kontrollwiederfindung Recovery of controls	Ceveron® TGA Control-High (nM Peak Thrombin) LOT 1Y51C00	100 - 163	129.7 OK
	Ceveron® TGA Control-Low (nM Peak Thrombin) LOT 2751C00	59 - 99	80.2 OK
Etiketten und Barcode Überprüfung Label and bar code verification		OK	

Durchführung von
Prepared by

Fikret NASUFI

2 5. MRZ. 2025

Datum / Date:

Freigabe von
Approved by

Michael KAFKA

2 5. MRZ. 2025

Datum / Date: