

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQB300AZ
---	---	----------

Ceveron TGA RB Kit 180T

REF	5006011-RUO
LOT	B353C00.02
	2026-03-31

Inhalt Content	Reagenz Reagent	LOT	
3x1 mL	Ceveron TGA RB	7T41C	2026-09-30
3x1.5 mL	Ceveron TGA BUF	9Q41C	2026-09-30
3x3 mL	Ceveron TGA SUB	8E41C00	2026-03-31
3x1 mL	Ceveron TGA Control High	1Y54C00	2026-05-31
3x1 mL	Ceveron TGA Control Low	2752C00	2026-04-30
1x25 mL	Ca-Chloridlösung 25mmol/L	9C41CB0	2029-03-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Kontrollwiederfindung Recovery of controls	Ceveron TGA Control-High (nM Peak Thrombin) LOT 1Y54C00	91 - 149	121	OK
	Ceveron TGA Control-Low (nM Peak Thrombin) LOT 2752C00	50 - 90	70	OK
Etiketten und Barcode Überprüfung Label and bar code verification			OK	

Durchführung von
Prepared by

Fikret NASUFI *Fikret Nasufi*

16. JUNI 2025
Datum / Date:

Freigabe von
Approved by

Lucas HOLTMAN
Lucas HOLTMAN

16. JUNI 2025
Datum / Date: