

# CERTIFICATE OF ANALYSIS

## Ceveron TGA RB Kit



Temp.: TF0327.07

Attachment 19      DQBx00.02

B362C00.02

REF	5006011-RUO
LOT	B362C00.02
	2027-10-31

Content	volume	Reagent	LOT	
3x	1mL	Ceveron TGA RB	7T53C	2027-11-30
3x	1.5 mL	Ceveron TGA BUF	9Q51C	2027-10-31
3x	1 mL	Ceveron TGA Control High	1Y61C00	2028-03-31
3x	1 mL	Ceveron TGA Control Low	2761C00	2028-03-31
3x	3mL	Ceveron TGA SUB	8E61C00	2028-03-31
1x	25mL	Ca-Chloridlösung 25 mmol/L	9C51CB0	2030-09-30

Analytical Procedures		Requirement	Result	
visual inspection	Content	according to packaging SOP	confirmed	OK
	Vial label	according to packaging SOP	confirmed	OK
	Vial barcode	according to packaging SOP	confirmed	OK
	Box label	according to packaging SOP	confirmed	OK
	Box barcode	according to packaging SOP	confirmed	OK
	Package insert	present	confirmed	OK
	Batch table	present	confirmed	OK
Recovery of controls	<b>Ceveron TGA Control High</b> LOT 1Y61C00	215 - 358 nM	346 nM	OK
	<b>Ceveron TGA Control Low</b> LOT 2761C00	91 - 165 nM	154 nM	OK

Prepared by

Fikret NASUFI *Fikret Nasufi*

15. APR. 2026

Date

Approved by

*Maylis* RAUSCHER

15. APR. 2026

Date