

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQB500AZ
---	---	----------

Ceveron® TGA RC High Kit 180T

REF 5006015-RUO

LOT B543C00.02

 2025-08-31

Inhalt Content	Reagenz Reagent	LOT	
3x1 mL	Ceveron® TGA RC High	7Q41C	2026-03-31
3x1.5 mL	Ceveron® TGA BUF	9Q41C	2026-09-30
3x3 mL	Ceveron® TGA SUB	8E41C00	2026-03-31
3x1 mL	Ceveron® TGA Control High	1Y42C00	2025-08-31
3x1 mL	Ceveron® TGA Control Low	2742C00	2025-08-31
1x25 mL	Ca-Chloridlösung 25mmol/L	9C41CB0	2029-03-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Kontrollwiederfindung Recovery of controls	Ceveron® TGA Control-High (nM Peak Thrombin) LOT 1Y42C00	352 - 586	484	OK
	Ceveron® TGA Control-Low (nM Peak Thrombin) LOT 2742C00	204 - 340	227	OK
Etiketten und Barcode Überprüfung Label and bar code verification			OK	

Durchführung von
Prepared by

Fikret NASUFI 

21. OKT. 2024

Datum / Date:

Freigabe von
Approved by

 L. HOLVICH

21. OKT. 2024

Datum / Date: