

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQB500AZ
---	---	----------

Ceveron TGA RC High Kit 180T

REF 5006015-RUO

LOT B555C00.02

 2026-09-30

Inhalt Content	Reagenz Reagent	LOT	
3x1 mL	Ceveron TGA RC High	7Q52C	2027-07-31
3x1.5 mL	Ceveron TGA BUF	9Q41C	2026-09-30
3x3 mL	Ceveron TGA SUB	8E52C00	2027-08-31
3x1 mL	Ceveron TGA Control High	1Y55C00	2027-08-31
3x1 mL	Ceveron TGA Control Low	2753C00	2027-08-31
1x25 mL	Ca-Chloridlösung 25mmol/L	9C41CB0	2029-03-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Kontrollwiederfindung Recovery of controls	Ceveron TGA Control-High (nM Peak Thrombin) LOT 1Y55C00	463 - 665	588	OK
	Ceveron TGA Control-Low (nM Peak Thrombin) LOT 2753C00	258 - 430	339	OK
Etiketten und Barcode Überprüfung Label and bar code verification			OK	

Durchführung von
Prepared by

Fikret NASUFI *Fikret Nasufl*

12. SEP. 2025

Datum / Date:

Freigabe von
Approved by

Lucas MOLOVICZ *Lucas*

12. SEP. 2025

Datum / Date: