



ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS

DQB500AZ

Ceveron TGA RC High Kit 180T

REF 5006015-RUO

LOT B554C00.02

2026-04-30

Inhalt Content	Reagenz Reagent	LOT	
3x1 mL	Ceveron TGA RC High	7Q52C	2027-07-31
3x1.5 mL	Ceveron TGA BUF	9Q41C	2026-09-30
3x3 mL	Ceveron TGA SUB	8E52C00	2027-08-31
3x1 mL	Ceveron TGA Control High	1Y56C00	2026-05-31
3x1 mL	Ceveron TGA Control Low	2754C00	2026-04-30
1x25 mL	Ca-Chloridlösung 25mmol/L	9C41CB0	2029-03-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Kontrollwiederfindung Recovery of controls	Ceveron TGA Control-High (nM Peak Thrombin) LOT 1Y56C00	434 - 724	593	OK
	Ceveron TGA Control-Low (nM Peak Thrombin) LOT 2754C00	252- 420	342	OK
Etiketten und Barcode Überprüfung Label and bar code verification			OK	

Durchführung von
Prepared by

Fikret NASUFI

22. AUG. 2025
Datum / Date:

Freigabe von
Approved by

Lucas HOLOVICZ

22. AUG. 2025
Datum / Date: