

CERTIFICATE OF ANALYSIS

TECHNOTHROMBIN TGA RC Low 5x0.5 mL



Attachment 03 DQ7D00.02

7D51C00.01

Temp.: T70336.07

REF 5006212
 LOT 7D51C00.01
 2027-03-31

Analytical Procedures		Requirement		Result	
visual inspection	Cap color	turquoise		confirmed	OK
	Vial type	clear glass vial		confirmed	OK
	Vial label	according to packaging SOP		confirmed	OK
	Vial barcode	according to packaging SOP		confirmed	OK
	Box label	according to packaging SOP		confirmed	OK
	Box barcode	according to packaging SOP		confirmed	OK
	Package insert	present		confirmed	OK
	Batch table	present		confirmed	OK
	Appearance before reconstitution	white colored plug of lyophilized material		confirmed	OK
	Appearance after reconstitution	clear liquid		confirmed	OK
Recovery of controls	TGA Control High LOT 1Z53C00	312	- 520 nM	465 nM	OK
	TGA Control Low LOT 2852C00	121	- 299 nM	220 nM	OK

Prepared by

Fikret NASUFI *Fikret Nasufi*

08. JULI 2025

Date

Approved by

Lucas
 Lucas HOLOVICZ

08. JULI 2025

Date