

**CERTIFICATE OF ANALYSIS
TECHNOTHROMBIN TGA RC Low 5x0.5 mL**



Attachment 03 DQ7D00.02

Temp.: TF0326.07

7D31C01.01

REF	5006212
LOT	7D31C01.01
⌚	2026-02-28

Analytical Procedures		Requirement				Result		
visual inspection	Cap color	turquoise				confirmed	OK	
	Vial type	clear glass vial				confirmed	OK	
	Vial label	according to packaging SOP				confirmed	OK	
	Vial barcode	according to packaging SOP				confirmed	OK	
	Box label	according to packaging SOP				confirmed	OK	
	Box barcode	according to packaging SOP				confirmed	OK	
	Package insert	present				confirmed	OK	
	Batch table	present				confirmed	OK	
	Appearance before reconstitution	white colored plug of lyophilized material				confirmed	OK	
	Appearance after reconstitution	clear liquid				confirmed	OK	
Recovery of controls	TGA Control High LOT 1Z41C00	269	-	448	nM	399	nM	OK
	TGA Control Low LOT 2841C00	94	-	214	nM	193	nM	OK

Prepared by Fikret NASUFI *Fikret Nasufi*

20. AUG. 2024

Date

Approved by *Lucas HOLLOWICZ*

21. AUG. 2024

Date