

**CERTIFICATE OF ANALYSIS
TECHNOTHROMBIN TGA RC Low 5x0.5 mL**



Attachment 03 DQ7D00.02

Temp.: TP0326.07

7D51C00.02

REF	5006212
LOT	7D51C00.02
	2027-03-31

Analytical Procedures		Requirement	Result	
visual inspection	Cap color	turquoise	confirmed	OK
	Vial type	clear glass vial	confirmed	OK
	Vial label	according to packaging SOP	confirmed	OK
	Vial barcode	according to packaging SOP	confirmed	OK
	Box label	according to packaging SOP	confirmed	OK
	Box barcode	according to packaging SOP	confirmed	OK
	Package insert	present	confirmed	OK
	Batch table	present	confirmed	OK
	Appearance before reconstitution	white colored plug of lyophilized material	confirmed	OK
	Appearance after reconstitution	clear liquid	confirmed	OK
Recovery of controls	TGA Control High LOT 1Z54C00	379 - 631 nM	453 nM	OK
	TGA Control Low LOT 2853C00	140 - 234 nM	180 nM	OK

Prepared by	<u>Fikrol NASUFI <i>Fikrol Nasufi</i></u>	<u>18. DEZ. 2025</u> Date
Approved by	<u><i>L. C.</i> Lukas GRUBELNIG</u>	<u>18. DEZ. 2025</u> Date