

CERTIFICATE OF ANALYSIS
TECHNOTHROMBIN TGA RC High 5x0.5 mL



Attachment 03 DQ7F00.04
7F53C00.01

Temp.: TP0336.07

REF 5006214
LOT 7F53C00.01
⌚ 2027-02-28

Analytical Procedures		Requirement		Result	
visual inspection	Cap color	turquoise		confirmed	OK
	Vial type	clear glass vial		confirmed	OK
	Vial label	according to packaging SOP		confirmed	OK
	Vial barcode	according to packaging SOP		confirmed	OK
	Box label	according to packaging SOP		confirmed	OK
	Box barcode	according to packaging SOP		confirmed	OK
	Package insert	present		confirmed	OK
	Batch table	present		confirmed	OK
	Appearance before reconstitution	white colored plug of lyophilized material		confirmed	OK
	Appearance after reconstitution	clear liquid		confirmed	OK
Recovery of controls	TGA Control High				
	LOT 1Z53C00	422	- 704 nM	633 nM	OK
	TGA Control Low				
	LOT 2852C00	240	- 416 nM	368 nM	OK

Prepared by

Fikret NASUFI *Fikret Nasufi*

0 8. JULI 2025
Date

Approved by

Lucas HOLOVICZ *Lucas Holovicz*

0 8. JULI 2025
Date