

**CERTIFICATE OF ANALYSIS**  
**TECHNOTHROMBIN TGA RC High 5x0.5 mL**



Attachment 03      DQ7F00.04

7F53C00.01

Temp.: TP0326.07

REF      5006214  
 LOT      7F53C00.01  
      2027-02-28

Analytical Procedures		Requirement		Result	
visual inspection	Cap color	turquoise		confirmed	OK
	Vial type	clear glass vial		confirmed	OK
	Vial label	according to packaging SOP		confirmed	OK
	Vial barcode	according to packaging SOP		confirmed	OK
	Box label	according to packaging SOP		confirmed	OK
	Box barcode	according to packaging SOP		confirmed	OK
	Package insert	present		confirmed	OK
	Batch table	present		confirmed	OK
	Appearance before reconstitution	white colored plug of lyophilized material		confirmed	OK
	Appearance after reconstitution	clear liquid		confirmed	OK
Recovery of controls	<b>TGA Control High</b>				
	LOT 1Z53C00	422	- 704 nM	633 nM	OK
	<b>TGA Control Low</b>				
	LOT 2852C00	240	- 416 nM	368 nM	OK

Prepared by

Fikret NASUFI *Fikret Nasufi*

08. JULI 2025

Date

Approved by

*Lucas Holovicz*  
 LUCAS HOLOVICZ

08. JULI 2025

Date