

**CERTIFICATE OF ANALYSIS**  
**TECHNOTHROMBIN TGA RC High 5x0.5 mL**



Attachment 03      DQ7F00.04

Temp.: TPO326.07

7F53C00.02

REF	5006214
LOT	7F53C00.02
	2027-02-28

Analytical Procedures	Requirement	Result
visual inspection	Cap color	turquoise      confirmed      OK
	Vial type	clear glass vial      confirmed      OK
	Vial label	according to packaging SOP      confirmed      OK
	Vial barcode	according to packaging SOP      confirmed      OK
	Box label	according to packaging SOP      confirmed      OK
	Box barcode	according to packaging SOP      confirmed      OK
	Package insert	present      confirmed      OK
	Batch table	present      confirmed      OK
	Appearance before reconstitution	white colored plug of lyophilized material      confirmed      OK
	Appearance after reconstitution	clear liquid      confirmed      OK
Recovery of controls	<b>TGA Control High</b> LOT 1253C00	422      -      704      nM      633      nM      OK
	<b>TGA Control Low</b> LOT 2852C00	240      -      416      nM      368      nM      OK

Prepared by

Fikret NASUFI *Fikret Nasufi*

19. SEP. 2025

Date

Approved by

*Lucas*  
Lucas WOLOVICZ

19. SEP. 2025

Date