

CERTIFICATE OF ANALYSIS

TECHNOTHROMBIN TGA RC High 5x0.5 mL



Attachment 03 DQ7F00.04

Temp.: T0326.07

7F53C00.02

REF 5006214
LOT 7F53C00.02
2027-02-28

| Analytical Procedures | | Requirement | | Result | |
|-----------------------|--|--|----------|-----------|----|
| visual inspection | Cap color | turquoise | | confirmed | OK |
| | Vial type | clear glass vial | | confirmed | OK |
| | Vial label | according to packaging SOP | | confirmed | OK |
| | Vial barcode | according to packaging SOP | | confirmed | OK |
| | Box label | according to packaging SOP | | confirmed | OK |
| | Box barcode | according to packaging SOP | | confirmed | OK |
| | Package insert | present | | confirmed | OK |
| | Batch table | present | | confirmed | OK |
| | Appearance before reconstitution | white colored plug of lyophilized material | | confirmed | OK |
| | Appearance after reconstitution | clear liquid | | confirmed | OK |
| Recovery of controls | TGA Control High LOT 1Z53C00 | 422 | - 704 nM | 633 nM | OK |
| | TGA Control Low LOT 2852C00 | 240 | - 416 nM | 368 nM | OK |

Prepared by

Fikret NASUFI *Fikret Nasufi*

19. SEP. 2025

Date

Approved by

Lucas
Lucas WOŁOWICZ

19. SEP. 2025

Date